

Transducer Amplifier PC Board

P/N: 0307-1300-6

R-4 - Is the zero pot that sets the emission meter zero before turning on the filament

R-13 - Is the Full Scale adjustment that is a factory adjustment.

